Autoclave Maintenance and Run Log -- Year _____

Please Date and Initial

	JAN	FEB	MAR	APR	MAY	JUNE	JULY	AUG	SEPT	OCT	NOV	DEC
MONTHLY SPORE TESTING RESULTS												
MONTHLY CLEANING												
ANNUAL CALIBRATION												

Sterility is **EVENT related**, not time related; Pack is considers sterile unless and event causes contamination (example: punctured, torn, cracked packs = unsterile; evidence of water damage or yellowed packs = unsterile). Have process to routine evaluation of sterile packs
Log Process: Write date and Load # on Pack (if more than 1 load is run in the same day, write date and load #1 and then date and load #2 etc)

Date	Time	Load #	Item(s) Autoclaved	Temperature (250-254 Degrees)	Steam Pressure (15-17 psi)	Duration of Run (30 minutes)	Person Responsible