

Dexa Scanner

Please note: The CA Radiologic Health Branch (RHB) has additional requirements

Ref: CCR Title 17 Sections 30111, 30305, 30404)

No lead apron or shield should be used because it would interfere with the test. The amount of radiation is so low that it is highly unlikely that a dosimeter would not move at all.

Providers need to have:

- Copy of Title 17
- Caution X-Ray in use sign on/near door
- Statement that provider has registered the scanner
 - Available on the Department of Radiology's web site
- Copy of a request for an inspection, or a certificate showing yearly inspections of scanner
- Operator of Dexa-Scan needs his/her certificate of completion of Dexa-Scan training
- Dosimeter for operator of scanner